

ABX Minotrol CRP



LOT

MC 431

CONTROL



(Exp.) **2024-07-05**

(YYYY - MM - DD)

Rev 2

		Minilysebio																
PARAMETRES PARAMETERS	UNITES UNITS	CONTROL				1	CONTROL				2	CONTROL				3	TOLERANCES TOLERANCE	
		MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE	MICROS				TOLERANCES TOLERANCE		
				60 8P 60 18P					60 8P 60 18P					60 8P 60 18P				
GB WBC	10 ³ /mm ³ ;10 ⁹ /l			2.1		± 0.4			7.8			± 1.0			20.8			± 2.2
GR RBC	10 ⁶ /mm ³ ;10 ¹² /l			2.24		± 0.15			4.52			± 0.18			5.57			± 0.25
HB HGB	g/dl			6.0		± 0.4			13.3			± 0.5			17.9			± 0.7
	g/l			60		± 4			133			± 5			179			± 7
	mmol/l			3.73		± 0.25			8.26			± 0.31			11.12			± 0.44
HT HCT	%			16.8		± 2.0			37.1			± 2.5			50.1			± 3.0
	l/l			0.168		± 0.020			0.371			± 0.025			0.501			± 0.030
VGM MCV	µm ³ ; fl			75		± 4			82			± 4			90			± 4
TGMH MCH	pg			26.8		± 2.4			29.4			± 2.4			32.1			± 2.8
	fmol			1.66		± 0.15			1.83			± 0.15			2.00			± 0.18
CCMH MCHC	g/dl			35.7		± 3.0			35.9			± 3.0			35.7			± 3.0
	g/l			357		± 30			359			± 30			357			± 30
	mmol/l			22.18		± 1.86			22.28			± 1.86			22.17			± 1.86
IDR RDW	%			14.5		± 3.0			14.5			± 3.0			14.0			± 3.0
PLAQ. PLTS	10 ³ /mm ³ ;10 ⁹ /l			80		± 20			274			± 40			536			± 60
VPM MPV	µm ³ ; fl			8.8		± 2.0			8.0			± 2.0			8.1			± 2.0
LYMPHO	#			1.4		± 0.2			2.7			± 0.5			3.5			± 1.1
	%			66.5		± 8.0			34.0			± 6.0			17.0			± 5.0
MONO	#			0.2		± 0.2			1.2			± 1.1			1.7			± 1.3
	%			8.0		± 7.0			15.3			± 13.8			8.0			± 6.0
GRANULO	#			0.5		± 0.2			4.0			± 1.0			15.6			± 2.1
	%			25.5		± 7.0			50.7			± 10.0			75.0			± 10.0
CRP	mg/dl			N/A		± 0.30			N/A			± 0.70			N/A			± 1.00
	mg/l			N/A		± 3.0			N/A			± 7.0			N/A			± 10.0

Ref: TEMP-1070 Rev.42 1300027267-A